

Inogen

Inogen One Oxygen Concentrator

March 06, 2006

Report No. INOG0009

Report Prepared By



www.nwemc.com
1-888-EMI-CERT


© 2006 Northwest EMC, Inc

EMC Test Report

Certificate of Test
Issue Date: March 06, 2006
Inogen
Model: Inogen One Oxygen Concentrator

| Emissions | | | | |
|---------------------|------------------------------|-------------------------------|-------------------------------------|--------------------------|
| Test Description | Specification | Test Method | Pass | Fail |
| Radiated Emissions | EN 60601-1-2 Class B:2001 | CISPR 11:2004-06 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Radiated Emissions | MIL-STD-461E:1999 | RE101 all Army:1999 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Radiated Emissions | RTCA DO-160D Category M:1997 | RTCA DO-160D:1997 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Conducted Emissions | EN 60601-1-2 Class B:2001 | CISPR 11:2004-06 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Conducted Emissions | FCC 15.107 Class B:2005-10 | ANSI C63.4:2003 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| Conducted Emissions | RTCA DO-160D Category M:1997 | RTCA DO-160D Power Leads:1997 | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

Modifications made to the product
See the Modifications section of this report

Approved By:

 Dean Ghizzone, President

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America.

Product compliance is the responsibility of the client, therefore the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. This Report may only be duplicated in its entirety. The results of this test pertain only to the sample(s) tested, the specific description is noted in each of the individual sections of the test report supporting this certificate of test.