



Inova Labs

Protocol Report

LifeChoice Activox {Gazelle} RTCA EMC Test Results

Document Number: PROT02012-009

REVISION HISTORY			
DATE	REVISION	SECTION	CHANGES
8/3/12	00	All	Original Release

NAME	TITLE	DEPARTMENT	AUTHOR OR APPROVER	SIGNATURE
Dragan Nebragic	VP	Engineering	Approver	 8/3/12
Mike Coppo	Mgr	Materials	Author	 8/3/12
Larry Swinney <i>Ready Williams for LS</i>	Mgr	Production	Approver	 8-3-12
Chris Gilbert	Mgr	Engineering	Approver	 8/3/12

Page 2 of 2	Inova Labs	Effective Date: 8/3/12
Protocol Report, LifeChoice Activox (Gazelle) RTCA EMC Test Results		Number: PROT02012..Q09

1. Summary

This protocol report presents test results obtained from Professional Testing, EMI, Inc. for RTCA/D0-160G EMC/EMI tests. Testing was conducted in 2012 on the LifeChoice Activox portable oxygen concentrator (POC). All emissions measured (results) were within limits, PASS.

2. Scope

This test applies to the LifeChoice Activox POC (model #XYC100) manufactured by Inova Labs of Austin, Texas.

3. Reference Documents

RTCA/D0-160G, RTCA, INC., December 8, 2010, Environmental Conditions and Test Procedures for Airborne Equipment

4. Results

Both conducted and radiated emissions were evaluated by Professional Testing, EMI, Inc. (PTI) per RTCA/D0-160G in July-August, 2012. All emissions measured (results) were within limits, PASS. See Attachments #1 for a copies of the data reports.

5. Attachments

Attachment #1 PTI Data Reports