

TECH TIP

ASR-11 Calibration of Step Attenuators and Test Equipment Not Listed In The Sypris Calibration List

Problem: Not all test equipment and attenuators are on contract for calibration by Sypris Test & Measurement. This Tech Tip provides direction for the calibration of test equipment items not covered by the Sypris contract.

Background: At one time FAA test equipment was calibrated by the FAA depot test equipment calibration lab. To expedite calibrations much of this work was outsourced to civilian contractor. Currently Sypris Test & Measurement is performing the calibration service. Sypris only calibrates items covered by their contract.

Recommendation: For test equipment items not on the Sypris calibration list. Process as you would any LIS order. Send in the out of calibration equipment as an R & R LIS order, justified as out of calibration “**Needs Cal Only**”. This is the process to calibrate step attenuators and other test equipment items not covered by the Sypris contract.