Inogen

Inogen One G-3

Report No. INOG0018



Report Prepared By Northwest EMC Inc.

NORTHWEST EMC – (888) 364-2378 – www.nwemc.com

California – Minnesota – Oregon – New York – Washington

NORTHWEST EMC

22975 NW Evergreen Parkway
Suite 400
Hillsboro, Oregon 97124

Certificate of Test

Last Date of Test: September 29, 2011
Inogen
Model: Inogen One G-3

Emissions

Test Description	Specification	Test Method	Pass/Fail
Radiated Emissions	RTCA DO-160F:2007, Category M	RTCA DO-160F:2007	**Pass**
Conducted Emissions	RTCA DO-160F:2007, Category M	RTCA DO-160F:2007	**Pass**

Deviations From Test Standards

None

Approved By:

Victor Ratinoff, Operations Manager

This report must not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the federal government of the United States of America.
Product compliance is the responsibility of the client, therefore the tests and equipment modes of operation represented in this report were agreed upon by the client, prior to testing. This Report may only be duplicated in its entirety. The results of this test pertain only to the sample(s) tested. The specific description is noted in each of the individual sections of the test report supporting this certificate of test.

NORTHWEST EMC

Revision History

Revision Number	Description	Date	Page Number
00	None		