VERIFICATION

May 29, 2006 Our File No.: AIR062-RTCA

AirSep Corporation
260 Creekside Drive
Buffalo, New York
USA, 14228-2075

NOT TRANSFERABLE

This Verification Certificate is hereby issued to the named GRANTEE and is VALID ONLY for the equipment identified hereon for use under the rules and regulations listed below:

APPLICANT NAME:	**AirSep Corporation**
PRODUCT UNDER TEST:	**FreeStyle Portable Oxygen Concentrator** **Configuration # 1: Internal Battery** **Configuration # 2: External Battery Package** **Configuration # 3: Elpac AC Power Supply**
MODEL NO.:	**FreeStyle**
APPLICABLE STANDARD:	**Radio Technical Commission for Aeronautics, RTCA/DO-160D, Section 21: Emission and Radio Frequency Energy**
EQUIPMENT CATEGORY:	**M**
EQUIPMENT TYPE:	**Airborne Equipment**

Note(s):

i. *Test methods employed conform to the following General Test Procedures:*
   - Radio Technical Commission for Aeronautics, RTCA/DO-160D, Section 21: Emission and Radio Frequency Energy

ii. *See attached report, UltraTech's File No.:* AIR062-RTCA, *dated* May 29, 2006 *for details and conditions of Verification Compliance.*

- *Assessed by ITI (UK) Competent Body, NVLAP (USA) Accreditation Body & ACA/AUSTEL (Australia)*
- *Recognized/Listed by FCC (USA)*
- *Accredited by Industry Canada under ACC-LAB (Europe/Canada MRA)*
- *All test results contained in this engineering test report are traceable to National Institute of Standards and Technology (NIST)*

Approved by: Tri M. Luu, P.Eng.
V.P. – Engineering

UltraTech

3000 Bristol Circle, Oakville, Ontario, Canada L6H 6G4
Telephone (905) 829-1570 Facsimile (905) 829-8050
Website: www.ultratech-labs.com, Email: vhk.ultratech@sympatico.ca