



EMC TEST REPORT

Date: July 9, 2009

Test: RTCA/DO-160E Section 21, Category M

Product Name: LifeChoice Portable Oxygen Concentrator by Inova Labs

Manufacturer: International Biophysics Corporation

World Wide Distributor: Inova Labs, LLC

Test Report Number: LIF-RTCA-09082008

Test Completion Date: September 8, 2008

Prepared by: Reginald Gould

Reviewed/Approved by:

Allan Westersten
Allan Westersten

This report is for the exclusive use of Dienco's Client and is provided pursuant to the agreement between Dienco and its Client. Dienco's responsibility and liability are limited to the terms and conditions of the agreement. Dienco assumes no liability to any party other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the client is authorized to copy or distribute this report and then only in its entirety. Any use of the Dienco name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Dienco. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product or service is or has been under a Dienco certification program. This report must not be used to claim product endorsement by A21.A, NIST nor any other agency of the U.S. Government.



**TEST SUMMARY
FOR
LIFECHOICE POC**

Test Method:

Test methods utilized on the EUT and the corresponding test results are shown below.

EUT: LifeChoice Portable Oxygen Concentrator.

Section	Method	Results
21.4	RTCA-160E Section 21, Category M, Conducted RF Emissions	Complies
21.5	RTCA-160E Section 21, Category M, Radiated RF Emissions	Complies

END OF REPORT