VERIFICATION

n n n n n



Our File No.: AIR062-RTCA

May 29, 2006

## **AirSep Corporation**

260 Creekside Drive Buffalo, New York USA, 14228-2075

## NOT TRANSFERABLE

This Verification Certificate is hereby issued to the named GRANTEE and is VALID ONLY for the equipment identified hereon for use under the rules and regulations listed below:

APPLICANT NAME:	AirSep Corporation
PRODUCT UNDER TEST:	FreeStyle Portable Oxygen Concentrator Configuration # 1: Internal Battery Configuration # 2: External Battery Package Configuration # 3: Elpac AC Power Supply
MODEL NO.:	FreeStyle
APPLICABLE STANDARD:	Radio Technical Commission for Aeronautics, RTCA/DO- 160D, Section 21: Emission and Radio Frequency Energy
EQUIPMENT CATEGORY:	Μ
EQUIPMENT TYPE:	Airborne Equipment

## Note(s):

i

Test methods employed conform to the following General Test Procedures:

- Radio Technical Commission for Aeronautics, RTCA/DO-160D, Section 21: Emission and Radio Frequency Energy
- *ii.* See attached report, UltraTech's File No.: AIR062-RTCA, dated May 29, 2006 for details and conditions of Verification Compliance.
- Assessed by ITI (UK) Competent Body, NVLAP (USA) Accreditation Body & ACA/AUSTEL (Australia)
- Recognized/Listed by FCC (USA)
- Accredited by Industry Canada under ACC-LAB (Europe/Canada MRA)
- All test results contained in this engineering test report are traceable to National Institute of Standards and Technology (NIST)

Approved by: Tri M. Luu, P.Eng. V.P. – Engineering

UltraTech

3000 Bristol Circle, Oakville, Ontario, Canada L6H 6G4 Telephone (905) 829-1570 Facsimile (905) 829-8050 Website: www.ultratech-labs.com, Email: vhk.ultratech@sympatico.ca