





# Inova Labs

## Protocol Report

### LifeChoice Activox {Gazelle} RTCA EMC Test Results

Document Number: PROT02012-009

REVISION HISTORY			
DATE	REVISION	SECTION	CHANGES
8/3/12	00	All	Original Release

NAME	TITLE	DEPARTMENT	AUTHOR OR APPROVER	SIGNATURE
Dragan Nebragic	VP	Engineering	Approver	 8/3/12
Mike Coppo	Mgr	Materials	Author	 8/3/12
Larry Swinney <i>Ready Williams for LS</i>	Mgr	Production	Approver	 8-3-12
Chris Gilbert	Mgr	Engineering	Approver	 8/3/12

Page 2 of 2	Inova Labs	Effective Date: 8/3/12
Protocol Report, LifeChoice Activox (Gazelle) RTCA EMC Test Results		Number: PROT02012..Q09

## 1. Summary

This protocol report presents test results obtained from Professional Testing, EMI, Inc. for RTCA/D0-160G EMC/EMI tests. Testing was conducted in 2012 on the LifeChoice Activox portable oxygen concentrator (POC). All emissions measured (results) were within limits, PASS.

## 2. Scope

This test applies to the LifeChoice Activox POC (model #XYC100) manufactured by Inova Labs of Austin, Texas.

## 3. Reference Documents

RTCA/D0-160G, RTCA, INC., December 8, 2010, Environmental Conditions and Test Procedures for Airborne Equipment

## 4. Results

Both conducted and radiated emissions were evaluated by Professional Testing, EMI, Inc. (PTI) per RTCA/D0-160G in July-August, 2012. All emissions measured (results) were within limits, PASS. See Attachments #1 for a copies of the data reports.

## 5. Attachments

Attachment #1 PTI Data Reports